

**13th CONFERENCE for
YOUNG SCIENTISTS in CERAMICS**

**PROGRAMME
and
BOOK OF ABSTRACTS**

**October 16-19, 2019
Novi Sad, Serbia**

Programme and Book of Abstracts of The 13th Conference for Young Scientists in Ceramics (CYSC-2019) publishes abstracts from the field of ceramics, which are presented at traditional international Conference for Young Scientists in Ceramics.

Editors-in-Chief

Prof. Dr. Vladimir V. Srdić

Publisher

Faculty of Technology, University of Novi Sad
Bul. cara Lazara 1, 21000 Novi Sad, Serbia

For Publisher

Prof. Dr. Biljana Pajin

Printing layout

Vladimir V. Srdić, Marija Milanović, Ivan Stijepović

Press

SLUŽBENI GLASNIK, Beograd

CIP – Каталогизација у публикацији
Библиотека Матице српске, Нови Сад

666.3/.7(048.3)

CONFERENCE for Young Scientists in Ceramics (13 ; 2019 ; Novi Sad)

Programme and book of abstracts / 13th Conference for Young Scientists in Ceramics (CYSC-2017), October 16-19, 2019, Novi Sad ; [editor-in-chief Vladimir V. Srdić]. - Novi Sad : Faculty of Technology, 2019 (Beograd : Službeni glasnik). - XX, 152 str. : ilustr. ; 24 cm

Tiraž 180. - Registar.

ISBN 978-86-6253-104-9

a) Керамика - Технологија - Апстракти
COBISS.SR-ID 331006727



The Book of Abstracts of the 13th Conference for Young Scientists in Ceramics
is licensed under a
[Creative Commons Attribution-NonCommercial-NoDerivatives 4.0 International License](https://creativecommons.org/licenses/by-nc-nd/4.0/)

OC-109

**IN SITU SYNTHESIS AND CHARACTERIZATION OF
HYDROXYAPATITE/TITANIUM OXIDE COATINGS DERIVED BY
ANODIZATION AND ANAPHORETIC DEPOSITION**

Marijana Pantović Pavlović^{1,2}, Miroslav Pavlović¹, Sanja Eraković¹, Tanja Barudžija³, Jasmina Stevanović^{1,2}, Nenad Ignjatović⁴, Vladimir Panić^{1,2}

¹*Institute of Chemistry, Technology and Metallurgy, University of Belgrade, Njegoševa 12, Belgrade, Serbia*

²*Centre of Excellence in Environmental Chemistry and Engineering - ICTM, University of Belgrade, Njegoševa 12, Belgrade, Serbia*

³*Vinča Institute of Nuclear Sciences, University of Belgrade, 12-14 Mike Petrovića Street, Belgrade, Serbia*

⁴*Institute of Technical Science of the Serbian Academy of Sciences and Arts, Knez Mihailova 35, Belgrade, Serbia*
e-mail: m.pantovic@ihm.bg.ac.rs

Titanium substrates were electrochemically treated by novel *in-situ* synthesis method to produce anaphoretic hydroxyapatite/titanium oxide (HAp/TiO₂) coating. This synthesis method was performed *via* anaphoretic deposition of hydroxyapatite and simultaneous anodization of Ti in order to produce highly adherent and strengthened composite coating, where the influence of electric potential, time, electrolyte concentration and pH value of the anodization process on titanium surface roughness and adhesion strength of composite HAp/TiO₂ coatings was examined. *In situ* synthesis was performed in different solutions containing different concentrations of NaOH, HAp and ethanol at different values of time and constant voltage. The effects of anodizing voltage on the morphology and bioactivity of the HAp coating and on the bonding strength between the HAp coating and the anodized substrates were investigated. A modified chemical precipitation method was used to prepare hydroxyapatite powder by the reaction of stoichiometric amount of calcium oxide and phosphoric acid. Prior to novel *in situ* synthesis of HAp/TiO₂ composite coatings, optimization of anodization process of titanium was performed. Anodization was executed under different electric potentials and different distances of counter electrodes from working electrodes, but all experiments had constant quantity of electric charge. Results indicated that highly ordered amorphous TiO₂ surface is formed on the Ti substrate after anodic oxidation. Characterization of titanium samples having rectangular contact surfaces of 10×10×0.89 mm included SEM/EDS analyses, X-ray diffraction analyses, AFM surface topography, morphology and roughness analyses and linear measurements of roughness. The prepared coatings were characterized by FE-SEM, XRD and EDS. Adhesion was investigated by ASTM D 3359–97 Test method B. Uniform and adherent HAp/TiO₂ composite coating on Ti was obtained. The results of simulated body fluid immersing experiments suggest that obtained HAp coating exhibits promising bioactivity. The obtained coating can be good candidate for bone implants due to improved adhesion.